
Return Loss bridge
Generator
In
50
50
Load
Under Test
Reference
50
External Detector
10n
100k
DC Out to
Multimeter
1N5711
100n
Transformer is 1:1